



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga  
 Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga  
 Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795 Email: [echemenvironmental@gmail.com](mailto:echemenvironmental@gmail.com)  
 Technical: 0917-164-2927 • 0917-164-2944

DOH Accreditation No.03-014-1820-LW-2 Valid until December 2020  
 DENR C.R.No. 074/2017 Valid until October 16, 2020

## TEST REPORT

Client: **Manila Water Philippine Ventures (MWPV)** Test Report No.: **TR-20C-133A**  
 Address: **Level 3 2nd Phase, DPO Bldg., UP Town Corporate Center, 216 Katipunan Ave., Diliman, Quezon City** Date and Time Received: **03 Mar 2020 @04:35PM**  
 Contact Person: **Ms. Florisza I. Faustino** Sampled By: **ECHEM**  
 COC Number: **2020-C0037** Report Date: **11 Mar 2020**

Sample ID: **S20-C03-025D** Sampling Date: **03 Mar 2020**  
 Sample Description: **N5** Sampling Time: **09:43 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:17AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:17AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:17AM	AHP

Sample ID: **S20-C03-026D** Sampling Date: **03 Mar 2020**  
 Sample Description: **SRCC Boston** Sampling Time: **09:50 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:20AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:20AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:20AM	AHP

Sample ID: **S20-C03-027D** Sampling Date: **03 Mar 2020**  
 Sample Description: **N4** Sampling Time: **09:54 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:23AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:23AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:23AM	AHP

Sample ID: **S20-C03-028D** Sampling Date: **03 Mar 2020**  
 Sample Description: **N7** Sampling Time: **10:01 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:28AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:26AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:26AM	AHP

Sample ID: **S20-C03-029D** Sampling Date: **03 Mar 2020**  
 Sample Description: **Mr. Gregorio Residence** Sampling Time: **10:06 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:29AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:29AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:29AM	AHP



Sample ID:	<b>S20-C03-030D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Reservoir 1</b>	Sampling Time:	<b>10:09 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:32AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:32AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:32AM	AHP

Sample ID:	<b>S20-C03-031D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N2</b>	Sampling Time:	<b>10:15 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:35AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:35AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:35AM	AHP

Sample ID:	<b>S20-C03-032D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N1</b>	Sampling Time:	<b>10:21 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:38AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:38AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:38AM	AHP

Sample ID:	<b>S20-C03-033D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N3</b>	Sampling Time:	<b>10:26 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:41AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:41AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:41AM	AHP

Sample ID:	<b>S20-C03-034D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 7</b>	Sampling Time:	<b>10:31 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:44AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:44AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:44AM	AHP

Sample ID:	<b>S20-C03-035D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>BNC Pawikan</b>	Sampling Time:	<b>10:38 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:47AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:47AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:47AM	AHP



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga

Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga

Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795

Email: [echemenvironmental@gmail.com](mailto:echemenvironmental@gmail.com)

Technical: 0917-164-2927 • 0917-164-2944

Sample ID: <b>S20-C03-036D</b>		Sampling Date: <b>03 Mar 2020</b>					
Sample Description: <b>Deepwell 6</b>		Sampling Time: <b>10:44 AM</b>					
Sample Matrix: <b>Drinking Water</b>		Sampling Location: <b>Anvaya Cove, Morong, Bataan</b>					
Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<i>E.coli</i>	Absent	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:50AM	AHP
Thermotolerant Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:50AM	AHP
Heterotrophic Plate Count (HPC)	<1	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
Total Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:50AM	AHP
Sample ID: <b>S20-C03-037D</b>		Sampling Date: <b>03 Mar 2020</b>					
Sample Description: <b>Deepwell 5</b>		Sampling Time: <b>10:50 AM</b>					
Sample Matrix: <b>Drinking Water</b>		Sampling Location: <b>Anvaya Cove, Morong, Bataan</b>					
Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<i>E.coli</i>	Absent	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:53AM	AHP
Thermotolerant Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:53AM	AHP
Heterotrophic Plate Count (HPC)	<1	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
Total Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:53AM	AHP
Sample ID: <b>S20-C03-038D</b>		Sampling Date: <b>03 Mar 2020</b>					
Sample Description: <b>Sports Center</b>		Sampling Time: <b>10:56 AM</b>					
Sample Matrix: <b>Drinking Water</b>		Sampling Location: <b>Anvaya Cove, Morong, Bataan</b>					
Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<i>E.coli</i>	Absent	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @09:56AM	AHP
Thermotolerant Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:53AM	AHP
Heterotrophic Plate Count (HPC)	<1	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
Total Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @09:56AM	AHP
Sample ID: <b>S20-C03-040D</b>		Sampling Date: <b>03 Mar 2020</b>					
Sample Description: <b>Villas</b>		Sampling Time: <b>11:04 AM</b>					
Sample Matrix: <b>Drinking Water</b>		Sampling Location: <b>Anvaya Cove, Morong, Bataan</b>					
Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<i>E.coli</i>	Absent	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:02AM	AHP
Thermotolerant Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:02AM	AHP
Heterotrophic Plate Count (HPC)	<1	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
Total Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:02AM	AHP
Sample ID: <b>S20-C03-041D</b>		Sampling Date: <b>03 Mar 2020</b>					
Sample Description: <b>BNC Backdoor</b>		Sampling Time: <b>11:09 AM</b>					
Sample Matrix: <b>Drinking Water</b>		Sampling Location: <b>Anvaya Cove, Morong, Bataan</b>					
Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<i>E.coli</i>	Absent	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:05AM	AHP
Thermotolerant Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:05AM	AHP
Heterotrophic Plate Count (HPC)	<1	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
Total Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:05AM	AHP
Sample ID: <b>S20-C03-042D</b>		Sampling Date: <b>03 Mar 2020</b>					
Sample Description: <b>Beechcroft</b>		Sampling Time: <b>11:15 AM</b>					
Sample Matrix: <b>Drinking Water</b>		Sampling Location: <b>Anvaya Cove, Morong, Bataan</b>					
Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<i>E.coli</i>	Absent	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:06AM	AHP
Thermotolerant Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:06AM	AHP
Heterotrophic Plate Count (HPC)	<1	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
Total Coliform	<1.1	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:06AM	AHP



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga

Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga

Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795

Email: [echemenvironmental@gmail.com](mailto:echemenvironmental@gmail.com)

Technical: 0917-164-2927 • 0917-164-2944

Sample ID:	<b>S20-C03-043D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Golf Kitchen</b>	Sampling Time:	<b>11:22 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:09AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:09AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:09AM	AHP

Sample ID:	<b>S20-C03-044D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Caddie</b>	Sampling Time:	<b>11:26 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:12AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:12AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:12AM	AHP

Sample ID:	<b>S20-C03-045D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 8</b>	Sampling Time:	<b>11:34 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:15AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:15AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:15AM	AHP

Sample ID:	<b>S20-C03-046D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 9</b>	Sampling Time:	<b>11:42 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:18AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:18AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:18AM	AHP

Sample ID:	<b>S20-C03-047D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Common Area N9</b>	Sampling Time:	<b>11:46 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:21AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:21AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:21AM	AHP

Sample ID:	<b>S20-C03-048D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 10</b>	Sampling Time:	<b>11:51 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:24AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:24AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:24AM	AHP



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga

Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga

Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795

Email: [echemenvironmental@gmail.com](mailto:echemenvironmental@gmail.com)

Technical: 0917-164-2927 • 0917-164-2944

Sample ID:	<b>S20-C03-049D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>SBV Condo</b>	Sampling Time:	<b>11:56 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:27AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:27AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:27AM	AHP

Sample ID:	<b>S20-C03-050D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Reservoir 2</b>	Sampling Time:	<b>12:01 PM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:29AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:29AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:29AM	AHP

Sample ID:	<b>S20-C03-051D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N6B</b>	Sampling Time:	<b>12:06 PM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:32AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:32AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:32AM	AHP

Sample ID:	<b>S20-C03-052D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 11</b>	Sampling Time:	<b>12:13 PM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Remarks	Method	Date Analyzed	Analyst
<b>E.coli</b>	<b>Absent</b>	MPN/100 mL	Absent	PASSED	MTFT Thermotolerant Coliform Test (EC Medium)	04 Mar 2020 @10:38AM	AHP
<b>Thermotolerant Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:35AM	AHP
<b>Heterotrophic Plate Count (HPC)</b>	<b>&lt;1</b>	CFU/mL	< 500	PASSED	Pour Plate	04 Mar 2020 @10:56AM	AHP
<b>Total Coliform</b>	<b>&lt;1.1</b>	MPN/100 mL	< 1.1	PASSED	Multiple Tube Fermentation Technique	04 Mar 2020 @10:35AM	AHP

### Notes & References:

Standard Methods for the Examination of Water and Wastewater, 23rd Edition.  
DOH Administrative Order No. 2017-0010, PNSDW 2017.

Certified by:

*Anita H. Pineda* 11 Mar 2020  
Anita H. Pineda, RMT  
Medical Technologist, PRC No. 52546

Approved by:

*Jennylyn G. Nerpio* 11 Mar 2020  
Jennylyn G. Nerpio, RCh  
Laboratory Manager, PRC No. 12226

*May Ann M. Estavillo* 11 Mar 2020  
May Ann M. Estavillo, RCh  
QA/QC Manager, PRC No. 8652

Test results written in this report are based on the samples as received by the laboratory. No alteration or modification of this test report shall be binding against Echem unless in writing and signed by a duly authorized representative of the company. This report shall not be reproduced, except in full, without the written permission from the laboratory.

Not valid without  
company seal



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga  
 Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga  
 Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795 Email: [ecchemenvironmental@gmail.com](mailto:ecchemenvironmental@gmail.com)  
 Technical: 0917-164-2927 • 0917-164-2944

DOH Accreditation No.03-014-1820-LW-2 Valid until December 2020  
 DENR C.R.No. 074/2017 Valid until October 16, 2020

## TEST REPORT

Client: **Manila Water Philippine Ventures (MWPV)** Test Report No.: **TR-20C-133B**  
 Address: **Level 3 2nd Phase, DPO Bldg., UP Town Corporate Center, 216 Katipunan Ave., Diliman, Quezon City** Date and Time Received: **03 Mar 2020 @04:35PM**  
 Contact Person: **Ms. Florisza I. Faustino** Sampled By: **ECHEM**  
 COC Number: **2020-C0037** Report Date: **16 Mar 2020**

Sample ID: **S20-C03-025D** Sampling Date: **03 Mar 2020**  
 Sample Description: **N5** Sampling Time: **09:43 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID: **S20-C03-026D** Sampling Date: **03 Mar 2020**  
 Sample Description: **SRCC Boston** Sampling Time: **09:50 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID: **S20-C03-027D** Sampling Date: **03 Mar 2020**  
 Sample Description: **N4** Sampling Time: **09:54 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID: **S20-C03-028D** Sampling Date: **03 Mar 2020**  
 Sample Description: **N7** Sampling Time: **10:01 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID: **S20-C03-029D** Sampling Date: **03 Mar 2020**  
 Sample Description: **Mr. Gregorio Residence** Sampling Time: **10:06 AM**  
 Sample Matrix: **Drinking Water** Sampling Location: **Anvaya Cove, Morong, Bataan**

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga

Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga

Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795

Email: [ecchemenvironmental@gmail.com](mailto:ecchemenvironmental@gmail.com)

Technical: 0917-164-2927 • 0917-164-2944

Sample ID:	<b>S20-C03-030D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Reservoir 1</b>	Sampling Time:	<b>10:09 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-031D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N2</b>	Sampling Time:	<b>10:15 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-032D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N1</b>	Sampling Time:	<b>10:21 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-033D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N3</b>	Sampling Time:	<b>10:26 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-034D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 7</b>	Sampling Time:	<b>10:31 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-035D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>BNC Pawikan</b>	Sampling Time:	<b>10:38 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-036D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 6</b>	Sampling Time:	<b>10:44 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	0.15	mg/L	0.40	Persulfate	13 Mar 2020	SAG



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga

Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga

Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795

Email: [echemenvironmental@gmail.com](mailto:echemenvironmental@gmail.com)

Technical: 0917-164-2927 • 0917-164-2944

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
<b>Sample ID: S20-C03-037D</b>						
<b>Sample Description: Deepwell 5</b>						
<b>Sample Matrix: Drinking Water</b>						
<b>Sampling Date: 03 Mar 2020</b>						
<b>Sampling Time: 10:50 AM</b>						
<b>Sampling Location: Anvaya Cove, Morong, Bataan</b>						
Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	0.18	mg/L	0.40	Persulfate	13 Mar 2020	SAG
<b>Sample ID: S20-C03-038D</b>						
<b>Sample Description: Sports Center</b>						
<b>Sample Matrix: Drinking Water</b>						
<b>Sampling Date: 03 Mar 2020</b>						
<b>Sampling Time: 10:56 AM</b>						
<b>Sampling Location: Anvaya Cove, Morong, Bataan</b>						
Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG
<b>Sample ID: S20-C03-039D</b>						
<b>Sample Description: Sports Center Kitchen</b>						
<b>Sample Matrix: Drinking Water</b>						
<b>Sampling Date: 03 Mar 2020</b>						
<b>Sampling Time: 10:59 AM</b>						
<b>Sampling Location: Anvaya Cove, Morong, Bataan</b>						
Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG
<b>Sample ID: S20-C03-040D</b>						
<b>Sample Description: Villas</b>						
<b>Sample Matrix: Drinking Water</b>						
<b>Sampling Date: 03 Mar 2020</b>						
<b>Sampling Time: 11:04 AM</b>						
<b>Sampling Location: Anvaya Cove, Morong, Bataan</b>						
Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG
<b>Sample ID: S20-C03-041D</b>						
<b>Sample Description: BNC Backdoor</b>						
<b>Sample Matrix: Drinking Water</b>						
<b>Sampling Date: 03 Mar 2020</b>						
<b>Sampling Time: 11:09 AM</b>						
<b>Sampling Location: Anvaya Cove, Morong, Bataan</b>						
Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG
<b>Sample ID: S20-C03-042D</b>						
<b>Sample Description: Beechcroft</b>						
<b>Sample Matrix: Drinking Water</b>						
<b>Sampling Date: 03 Mar 2020</b>						
<b>Sampling Time: 11:15 AM</b>						
<b>Sampling Location: Anvaya Cove, Morong, Bataan</b>						
Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG
<b>Sample ID: S20-C03-043D</b>						
<b>Sample Description: Golf Kitchen</b>						
<b>Sample Matrix: Drinking Water</b>						
<b>Sampling Date: 03 Mar 2020</b>						
<b>Sampling Time: 11:22 AM</b>						
<b>Sampling Location: Anvaya Cove, Morong, Bataan</b>						
Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga

Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga

Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795

Email: [echemenvironmental@gmail.com](mailto:echemenvironmental@gmail.com)

Technical: 0917-164-2927 • 0917-164-2944

Sample ID:	<b>S20-C03-044D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Caddie</b>	Sampling Time:	<b>11:26 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-045D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 8</b>	Sampling Time:	<b>11:34 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-046D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 9</b>	Sampling Time:	<b>11:42 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-047D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>NP</b>	Sampling Time:	<b>11:46 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-048D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 10</b>	Sampling Time:	<b>11:51 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-049D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>SBV Card O</b>	Sampling Time:	<b>11:56 AM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-050D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Reservoir 2</b>	Sampling Time:	<b>12:01 PM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG



# Echem ENVIRONMENTAL TESTING LABORATORY CORPORATION

Laboratory: 10 Quezon Road, San Isidro, San Simon, Pampanga

Satellite Office: G/F ASCORP Bldg., Mc Arthur Highway, Dolores, City of San Fernando, Pampanga

Sales: 0942-316-2538 • 0945-798-6962 | (045) 435-2795

Email: [echemenvironmental@gmail.com](mailto:echemenvironmental@gmail.com)

Technical: 0917-164-2927 • 0917-164-2944

Sample ID:	<b>S20-C03-051D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>N6B</b>	Sampling Time:	<b>12:06 PM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	<0.14	mg/L	0.40	Persulfate	13 Mar 2020	SAG

Sample ID:	<b>S20-C03-052D</b>	Sampling Date:	<b>03 Mar 2020</b>
Sample Description:	<b>Deepwell 11</b>	Sampling Time:	<b>12:13 PM</b>
Sample Matrix:	<b>Drinking Water</b>	Sampling Location:	<b>Anvaya Cove, Morong, Bataan</b>

Parameter	Result	Unit	Limit	Method	Date Analyzed	Analyst
Iron	<0.10	mg/L	1.0	Phenanthroline	11 Mar 2020	SAG
Manganese	0.34	mg/L	0.40	Persulfate	13 Mar 2020	SAG

### Notes & References:

Standard Methods for the Examination of Water and Wastewater, 23rd Edition.

DOH Administrative Order No. 2017-0010, PNSDW 2017.

#### Certified by:

  
**Jennilyn G. Nerpio, RCh**  
 Laboratory Manager, PRC No. 12226

#### Approved by:

  
**May Ann M. Estavillo, RCh**  
 QA/QC Manager, PRC No. 8652

Test results written in this report are based on the samples as received by the laboratory. No alteration or modification of this test report shall be binding against Echem unless in writing and signed by a duly authorized representative of the company. This report shall not be reproduced, except in full, without the written permission from the laboratory.

Not valid without  
company seal